


The conditions:

Column:	4.6*250mm, BIO-120-5-ODS 5um		
Solvent A :	0.1%Trifluoroacetic in 100% Acetonitrile		
Solvent B :	0.1%Trifluoroacetic in 100% Water		
Gradient:	A	B	
	0.01min	30%	70%
	15.00min	55%	45%
	15.10min	100%	0%
	20.0min	STOP	
Flow rate:	1.0ml/min		
Wavelength:	220nm		
Volume:	5ul		

The Map :



The data:

Rank	Time	Conc.	Area	Height
1	7.933	0.1362	73900	3993
2	8.065	0.1188	64423	13112
3	8.335	99.66	54055347	1903297
4	9.188	0.09134	49545	3533
Total		100	54243215	1923935